Assay of Creatine Kinase Isoenzyme MB in Serum with Time-Resolved Immunofluorometry

E. S. Lianidou,^{1,2} T. K. Christopoulos,¹ and E. P. Diamandis^{1,3,4}

We describe the first time-resolved immunofluorometric assay for creatine kinase (EC 2.7.3.2) isoenzyme MB (CK-MB) in serum. The assay is based on the formation of the complex: solid-phase anti-CK-MB–CK-MB–biotinylated anti-CK-BB– streptavidin–BCPDA–Eu³⁺, where anti-CK-MB and anti-CK-BB are monoclonal antibodies against the CK isoenzymes MB and BB, respectively, and BCPDA is the europium chelator 4,7-bis(chlorosulfophenyl)-1,10-phenanthroline-2,9-dicarboxylic acid. The solid-phase complex is fluorescent and is measured on the dry solid-phase (microtiter well) in a specially designed time-resolved fluorometer that uses laser excitation. The assay requires 25 μ L of serum and is not affected by the presence of either CK-MM (up to 5000 μ g/L) or CK-BB (up to 1000 μ g/L) in the sample. Precision and accuracy indices for the assay were satisfactory.

Additional Keyphrases: acute myocardial infarction · biotinstreptavidin · europium chelates

Measurement of serum creatine kinase (ATP:creatine V-phosphotransferase, EC 2.7.3.2) MB isoenzyme (CK-MB) [1) is still the most specific and sensitive indicator of nyocardial necrosis and the most reliable and critical aboratory test for the biochemical diagnosis of acute myoardial infarction (AMI) (2, 3).⁵ Thus, the specificity and sensitivity of laboratory tests for the determination of CK-MB are of critical importance (4).

Several methods for determining CK-MB in serum have been developed. Methods based on determining catalytic activity after separation of CK-MB from the CK-BB and CK-MM isoenzymes by electrophoresis or ion-exchange thromatography are generally semiquantitative and labor intensive (5). Immunological methods for CK-MB based on mmunoinhibition or immunoprecipitation of CK-MB and neasurement of catalytic activity of CK-MB are susceptible to interferences from CK-BB isoenzyme and macro-CK type 1 or mitochondrial CK (macro-CK type 2) (6, 7).

Recently developed two-site immunoassays for CK-MB are superior to other immunochemical methods in sensitivty and specificity. These "sandwich"-type assay configuraions use highly specific anti-CK-MM, anti-CK-BB, or inti-CK-MB monoclonal antibodies. Several highly specific two-site immunoassays for CK-MB have recently become commercially available, e.g., the Stratus two-site fluorometric radial partition immunoassay (Baxter Healthcare Corp., Miami, FL 33152) (8), the MagicLite chemiluminometric immunoassay (Ciba Corning Diagnostics Corp., E. Walpole, MA 02032) (9), the Icon QSR assay (Hybritech, San Diego, CA 92126) (10), and the IMx fluorometric assay (Abbott Diagnostics, Abbott Park, IL 60064). Newly developed assays for CK isoforms in serum purportedly are even more specific markers for AMI but are not yet widely used (11).

Among the possible problems of the immunometric-type CK-MB assays are the falsely positive interferences by heterophilic antibodies (e.g., anti-mouse IgG antibodies present in human sera), which have been reported for many two-site immunoassays that involve mouse monoclonal antibodies (12). False-positive results were also reported for the Stratus CK-MB assay because of the presence of a high-molecular-mass form of alkaline phosphatase (13), although this interference seems to have been eliminated (13).

Recently, highly sensitive two-site immunoassays based on time-resolved fluorescence measurements of europium have been developed (14). Here, we describe the first time-resolved immunofluorometric assay of CK-MB in serum. The assay is based on the formation of an immunocomplex involving CK-MB, a specific monoclonal antibody to CK-MB immobilized on a solid phase (microtiter well), and a biotinylated monoclonal antibody specific for the B subunit. The formed immunocomplex is quantified by adding a highly sensitive 4,7-bis(chlorosulfophenyl)-1,10phenanthroline-2,9-dicarboxylic acid (BCPDA)-labeled streptavidin-based macromolecular complex, which has been recently described (15), and by measuring the specific Eu³⁺ fluorescence in a time-resolved fluorometer.

Materials and Methods

Instrumentation

For time-resolved solid-phase fluorescence measurements we used the CyberFluor 615^{m} Time-Resolved Fluorometer/Immunoanalyzer (CyberFluor Inc., Toronto, Canada) with an excitation wavelength of 337.1 nm (nitrogen laser source) and an emission wavelength of $615 (\pm 5)$ nm (interference filter). Specially designed software provides automated data reduction. White opaque microtitration wells (assembled in 12 wells per strip, Immulon II) were products of Dynatech Labs, Alexandria, VA 22314.

Materials

The europium chelator, BCPDA, was synthesized as previously described (16). Streptavidin and bovine serum albumin (BSA) were purchased from Sigma Chemical Co., St. Louis, MO 63178. Sulfosuccinimidyl 6-(biotinamido) hexanoate (NHS-LC-biotin) was from Pierce Chemical Co., Rockford, IL 61105.

Creatine kinase isoenzymes. CK-MB from human heart tissue (purity \geq 99%), CK-MM from human skeletal muscle (purity \geq 98%), and CK-BB from human brain (purity

¹ Department of Clinical Biochemistry, Toronto Western Hospial, 399 Bathurst St., Toronto, Ontario M5T 2S8, Canada.

² Permanent address: Laboratory of Analytical Chemistry, Uniersity of Athens, 104 Solonos St., Athens, 106 80, Greece.

³ Department of Clinical Biochemistry, University of Toronto, 00 College St., Toronto, Ontario M5G 11.5, Canada.

⁴ Address correspondence to this author at the Toronto Western Iospital.

⁵ Nonstandard abbreviations: CK, creatine kinase; CK-MB, CK-B, and CK-MM, isoenzymes of CK; AMI, acute myocardial afarction; BCPDA, 4,7-bis(chlorosulfophenyl)-1,10-phenanthrone-2,9-dicarboxylic acid; SA, streptavidin; TG, bovine thyrogloblin; BSA, bovine serum albumin; and NHS-LC-biotin, sulfosucinimidyl 6-(biotinamido) hexanoate.

Received February 23, 1990; accepted July 3, 1990.

 \geq 95%) were purchased from Scripps Laboratories, San Diego, CA 92131. All purities were established by electrophoresis. All other chemicals were from Sigma Chemical Co.

CK-MB standards. CK-MB standards at concentrations of 0-100 μ g/L were prepared by appropriate dilutions of CK-MB in heat-inactivated and filtered normal pooled human serum. These standards were calibrated with the Stratus CK-MB assay and were kept at 4 °C when not in use and at -20 °C and for longer storage (>1 week).

Serum samples. Serum samples, obtained from 66 patients admitted to the coronary care unit of Toronto Western Hospital, were found positive for CK-MB by a screening immunoinhibition method (see below). Fifty-seven samples were obtained from a mixed population (Caucasian men and women) of apparently healthy subjects (blood donors), ages 20-40 years, for estimating the reference range.

Monoclonal antibodies. We purchased monoclonal antibodies for the B subunit of CK, anti-CK-BB (clone nos. 027-11288, 027-11612, 027-15712, 027-18812, 027-11267, 027-18432), as well as for the M subunit of CK, anti-CK-MM (clone nos. 027-17186, 027-18367, 027-12859, 027-11310, 027-14110), from OEM Concepts Inc., Toms River, NJ 08755. We have also purchased a specific monoclonal antibody for CK-MB, anti-CK-MB (lot no. 709107), and a specific monoclonal antibody for the B subunit of CK (lot no. 734916) from International Immunoassay Laboratories, Inc., Santa Clara, CA 95054. All monoclonal antibodies used in this study were purified from ascites fluid with an AffiGel protein A "MAPS" kit (Bio-Rad Laboratories, Richmond, CA 94804).

Procedures

Preparation of biotinylated antibodies. Monoclonal antibodies used as detection antibodies were biotinylated with sulfosuccinimidyl 6-(biotinamido) hexanoate (NHS-LC-biotin) as previously described (17). Working solutions of the biotinylated antibodies were prepared by diluting them to ~2.5–5 mg/L in 50 mmol/L Tris buffer, pH 7.40, containing 9 g of NaCl, 5 g of BSA, 0.5 g of NaN₃, 5 mL of normal mouse serum, 0.5 g of bovine γ -globulins, 0.5 mol of KCl, and 0.02 mL of Tween 40 per liter.

Preparation of labeled streptavidin. A streptavidin-based macromolecular complex, a very sensitive detection reagent, was formed by conjugating streptavidin with BCPDA-labeled thyroglobulin, as described elsewhere (18), and incubating the reaction mixture with Eu^{3+} as previously described (15). The concentration of streptavidin in this stock solution was 15 mg/L. Labeled streptavidin working solution was prepared by diluting the stock solution 50-fold in a 50 mmol/L Tris buffer, pH 7.20, containing, per liter, 40 g of bovine serum albumin, 9 g of NaCl, 0.1 g of sodium azide, and 40 μ mol of Eu^{3+} .

Coating of microtitration wells. The wells were coated overnight at room temperature with the coating solution (50 μ L/well), a 50 mmol/L Tris solution, pH 7.40, containing 10 μ g of coating antibody per milliliter. After coating, the wells were washed with a wash solution containing 9 g of NaCl, 0.5 mL of polyoxyethylene sorbitan monolaurate (Tween 20), and 0.5 g of NaN₃ per liter, and blocked for 30 min at room temperature with 200 μ L per well of a 50 mmol/L Tris solution, pH 7.40, containing 60 g of BSA and 0.5 g of sodium azide per liter. The wells were stored in the blocking solution and were washed just before use as described above.

ing to the manufacturers' recommendations. from 66 paboronto Westby a screening even samples lucasian men blood donors), e range. e range. e vert samples e vert samples

50 μ L of the labeled streptavidin working solution to eacl well. After 30-min incubation (or 15 min for the rapic procedure), wash the wells and then dry them with a forced-air plate dryer for 5 min at room temperature. The fluorescence of Eu³⁺ on the solid phase is measured with the CyberFluor immunoanalyzer. Data reduction is auto matic.

Comparison methods. We used four different methodolo

gies for comparison and other studies. As a screening

procedure, we used the immunoinhibition method, with reagents from Boehringer Mannheim Diagnostics, India

napolis, IN 46290, and a chemistry analyzer (Hitach

Model 717; Boehringer Mannheim). Sera found positive by

this method were further analyzed by three different two

site sandwich immunoassays specific for CK-MB, i.e., the

Stratus, the Icon QSR, and the IMx CK-MB assays, accord

Results

Selection of monoclonal antibodies. Thirteen monoclona antibodies were tested in various combinations in an initia effort to establish workable pairs. In all cases, we used a solid-phase antibody (immobilized on microtiter wells) and a biotinylated detection antibody. For coating, we tester the seven anti-CK-BB antibodies in combination with five anti-CK-MM detection antibodies (35 different pairs) and an anti-CK-MB detection antibody (seven pairs). Addition ally, we tested the anti-CK-MB antibody for coating and used either the anti-CK-BB (seven) or the anti-CK-MN antibodies for detection (total of 12 pairs). From the total 54 pairs tested, 24 pairs gave calibration curves. Furthe studies on cross-reactivity showed that only the pair o anti-CK-MB (coating antibody, lot no. 709107):anti-CK-BI (detection antibody, lot no. 734916) was free from botl positive and negative interference by CK-MM (up to 500) $\mu g/L$) or CK-BB (up to 1000 $\mu g/L$). To assess positive interference, we measured samples containing only CK MM or CK-BB; we assessed negative interference by check ing the percent reduction in the values found for a CK-MI sample initially measured as 20 μ g/L in the presence o

Table	1.	Interfere	nce S	tu	dies	for	the	Proposed	CK-MB
			Assa	y ((Pro	loco	ol A)		

Detection	Positive i	interference*	Negative interference ^b		
ng/weli	CK-MM	CK-BB	CK-MM	CK-BB	
500	ND (0) ^c	0.10 (0.01)	21.9 (109)	21.9 (109	
250	ND (0)	0.26 (0.03)	19.7 (98.5)	21.6 (108	
125	ND (0)	0.35 (0.04)	20.1 (100)	22.4 (112	

^a Expressed as $\mu g/L$ of CK-MB; CK-MM and CK-BB were added t CK-MB-free serum at concentrations of 5000 and 1000 $\mu g/L$, respectively. ^b Values shown are $\mu g/L$ of CK-BB measured for a sample original containing 20 μg of CK-MB per liter and then supplemented with CK-MM an

CK-BB at concentrations of 5000 and 1000 μ g/L, respectively. ^o Numbers in parentheses represent % cross-reactivity (positive interfe

ence studies) or % of expected values (100% when negative interference i zero). ND. not detected. CK-MM or CK-BB (Table 1). Thus, we finally selected anti-CK-MB as the solid-phase capture antibody and anti-CK-BB (lot no. 734916) as the biotinylated detection antibody for the development of this assay.

Assay optimization. The coating of wells was optimized for antibody concentration. An antibody concentration of 500 ng/50 μ L per well was chosen for maximum signal. We found that coating with 250, 125, or 60 ng per well yielded a signal of about 70%, 50%, and <40% of maximum, respectively.

With use of 500 ng per well of coating antibody, the assay was optimized for biotinylated-antibody concentration. An antibody concentration of 2.5 mg/L (125 ng/50 μ L per well) was chosen as optimum because it gave the lowest background signal, even though additional antibody (250 and 500 ng per well) gave practically the same fluorescence signals but a higher background. To study the effect of incubation time on the first step, we incubated the samples in the coated wells with biotinylated antibody for 15, 30, 60, and 120 min, but kept the streptavidin-binding incubation step constant at 30 min. Fluorescence signals were increased by increasing the incubation time up to 120 min but with no significant increase in assay detectability because nonspecific binding (zero standard signal) also increased. The same studies were performed with the streptavidin incubation step.

On the basis of these studies, we devised two assay protocols: a 15 min-15 min rapid protocol (A) and a 30 min-30 min normal protocol (B); the times refer to the first and second incubation steps, respectively. To minimize errors caused by variable incubation times between wells (delay effect), we first pipetted the detection antibody in the wells and then we pipetted all standards and patients' samples within 10 min, for both protocols. We also added 5 mL of mouse serum per liter of antibody diluent to avoid any possible falsely increased results for CK-MB, owing to rheumatoid-like factors and anti-mouse immunoglobulin antibodies in some human sera (12).

Assay characteristics. Typical standard curves for both assay protocols are shown in Figure 1. Both curves are



Fig. 1. Calibration curves for the proposed CK-MB assay for both protocols

□, rapid protocol (30 min); ◆, normal protocol (60 min). The fluorescence of he zero standard has been subtracted from all other fluorescence readings linear at concentrations of CK-MB $\leq 100 \ \mu g/L$. The detection limit, defined as the analyte concentration corresponding to the fluorescence reading of the zero standard +2 SD, was 0.46 and 0.36 $\mu g/L$ for the rapid and the normal protocol, respectively. The detection limit of the normal protocol showed no significant improvement because increasing the incubation times increased both the background and specific signal proportionally.

Precision. Within-run precision was determined by analyzing serum samples at three different concentrations, 12 times each, by both protocols. Day-to-day precision was assessed by analyzing another three serum samples 10 times over a period of 10 days (Table 2).

Dilution test. We evaluated the linearity of the method by diluting five different serum samples of known CK-MB concentrations (20–30 μ g/L) with heat-inactivated normal serum (already found negative for CK-MB by the Stratus method) and by reassaying them by both protocols. For all cases, linear dilution curves at dilution factors ranged from three- to 20-fold.

Analytical recovery. Serum samples supplemented with CK-MB at two different concentrations were prepared by adding known concentrations of exogenous CK-MB to four pooled serum samples with CK-MB values between 0.5 and 1.1 μ g/L. To assess analytical recovery, we analyzed the samples before and after the two additions of CK-MB (~18 and 40 μ g/L), with mean analytical recoveries of 82.8% \pm 5.7% and 91.9% \pm 4.6% for protocols A and B, respectively.

Reference interval. We analyzed 57 serum samples from apparently healthy blood donors, ages 20–40 years, by both proposed protocols. The reference interval, defined as the range of values between the 2.5th and 97.5th percentile, was 0–4.1 μ g/L. Measured values ranged from undetectable to 6.0 μ g/L.

Correlation with other methods. We compared the results from the proposed time-resolved immunofluorometric assay for CK-MB with those obtained by three commercially available immunoassays for CK-MB in serum, i.e., the Stratus, Icon QSR, and IMx CK-MB (Table 3). The proposed assay correlates well with the Stratus (r = 0.978) as well as with the Icon (r = 0.987) and IMx (r = 0.984). However, the slope of the regression equation is close to unity (0.84-0.90) only with the Stratus system and is significantly lower (0.58-0.71) for the Icon and IMx methods. This is expected in that the Icon assay reportedly gives higher values than the Stratus (19, 20). Although the IMx assay (21) has not been widely used, it seems to give higher readings than the Stratus assay.

	Table 2. Precis	ion of the	CK-MB Assay		
Sample					
	Moan (SD) CK-MB, μg/L	CV, %	CK-MB, μg/L	CV, %	
Within-rur	n (n = 12)				
1	7.4 (0.4)	5.4	8.4 (0.5)	6.0	
2	16.2 (1.0)	6.2	18.9 (1.1)	5.8	
3	39.6 (1.6)	4.0	45.2 (1.9)	4.2	
Day-to-da	y(n = 10)				
1	8.8 (1.1)	12.3	8.2 (0.9)	11.0	
2	26.0 (1.9)	7.3	22.8 (1.8)	7.9	
3	48.5 (4.0)	8.2	46.6 (4.6)	9.9	

Table 3.	Comparison of the	Proposed CK-MB As	say with Three	Two-Site Immunoassa	iys for CK-MB	
Comparison method (x)	Proposed method (y), protocol	CK-MB range, µg/L	Slope"	Intercept, µg/L	r	S _{y.}
Stratus, n = 66	Α	1.4-370	0.90	-0.29	0.978	17.
	В	0.9-309	0.84	3.75	0.962	20.
lcon, n = 43	Α	1.4-130	0.59	4.38	0.987	5.
	В	0. 9 –122	0.58	5.01	0.983	6.:
IMx, n = 43	Α	1.4-130	0.71	1.59	0.984	6.
	В	0. 9 –122	0.71	2.17	0.982	6.
* Linear regression.				,		

Discussion

The availability of specific monoclonal antibodies for the M and B subunits and MB isoenzyme of CK enabled the development of very specific and sensitive two-site immunoassays that measure CK-MB mass and use radiometric, fluorometric (8), chemiluminometric (9), or absorbance (10)detection systems. Correlation between measurements of CK-MB mass and activity is well established (22) and studies have successfully compared these two assay principles (23-25).

Time-resolved fluorescence has become an established analytical technique in the field of nonisotopic immunoassay and can achieve sensitivities equivalent to or better than those obtained with radiolabeled tracers (14). Here we describe a time-resolved immunofluorometric assay for CK-MB in serum. The assay is based on the formation of a complex involving CK-MB bound by a specific monoclonal antibody for CK-MB immobilized on the solid-phase and a biotinylated specific monoclonal antibody for the B subunit. The formed immunocomplex is quantified by a highly sensitive, recently described (15) streptavidin (SA)-based macromolecular complex of the approximate composition SA(TG)_{3.3}BCPDA₄₈₀, where TG is bovine thyroglobulin. This solid-phase complex is fluorescent and is measured on the surface of a microtiter well by a specially designed laserexcited, time-resolved fluorometer.

The proposed assay for CK-MB is highly sensitive, with a detection limit $\leq 0.4 \ \mu g/L$, equivalent to that reported for Stratus (0.4 μ g/L) but better than the reported detection limit of the MagicLite (0.8 μ g/L) and the Icon (0.6 μ g/L) assays (19). The reported detection limit of the IMx assay $(0.2 \ \mu g/L)$ is superior (21). Only 25 μL of sample is required for a single assay. The assay is free from positive and negative interferences by CK-MM (up to 5000 μ g/L) and CK-BB (up to 1000 μ g/L), and calibration curves are linear to CK-MB concentrations of 100 μ g/L. Serum dilution curves were linear over a wide range.

We have developed two different protocols, with total incubation times of either 30 or 60 min. Comparison studies with other commercially available immunoassays reveal good correlation although the absolute values differ, especially with the IMx and Icon assays. The reason for the discrepancies is unknown. We conclude that standardization must play a role because our assay is in close agreement with the Stratus assay, the method we used to assign values to our standards.

Our assay design, which incorporates a specific anti-CK-MB coating antibody and a specific anti-CK-B subunit antibody, is theoretically the most preferable. Assays that use anti-CK-B subunit coating antibody (the Icon assay) may suffer negative interference if high concentrations of CK-BB are present in serum. Also, use of anti-CK-I subunit detection antibody is preferable to the use o anti-CK-M subunit detection antibody (as used in the IM: assay) because of better specificity. The presence of high amounts of CK-MM in serum may cause problems if the coating antibody is not highly specific for CK-MB or if ther is nonspecific binding of CK-MM.

References

1. Lang H, Würzburg U. Creatine kinase, an enzyme of man forms [Review]. Clin Chem 1982;28:1439-47.

2. Moss DW, Henderson AR, Kachmar JF. Enzymes. In: Tiet NW, ed. Textbook of clinical chemistry. Philadelphia: Saunders 1986:619-763.

3. Lott JA, Stang JM. Serum enzymes and isoenzymes in the diagnosis and differential diagnosis of myocardial ischemia and necrosis [Review]. Clin Chem 1980;26:1241-50.

4. Kallner A, Sylven C, Brodin U, Loogna E, Svenhamn K. Early diagnosis of acute myocardial infarction. A comparison between chemical predictors. Scand J Clin Lab Invest 1989;49:633-9.

5. Koch TR, Mehta UJ, Nipper HC. Clinical and analytical eval uation of kits for measurements of creatine kinase isoenzyme MB Clin Chem 1986;32:186-91.

6. Stein W, Bohner J, Steinhart R, Eggstein M. Macro creating kinase: determination and differentiation of two types by their activation energies. Clin Chem 1982;28:19-24.

7. Wu AH, Bowers GN. Evaluation and comparison of immunoin hibition and immunoprecipitation methods for differentiating MI from BB and macroforms of creatine kinase isoenzymes in patient and healthy individuals. Clin Chem 1982;28:2017-21

8. Goodnow T, LeBlanc M, Koshi K, et al. Radial partition immunoassay: quantitating CK-MB levels in human serum [Ab stract]. Clin Chem 1987;33:988.

9. Piran U, Kohn DW, Uretsky LS, et al. Immunochemilumino metric assay of creatine kinase MB with a monoclonal antibody to the MB isoenzyme. Clin Chem 1987;33:1517-20.

10. Wu AHB, Gornet TG. Evaluation of an immunoenzymometric assay for CK-MB using the Icon format [Abstract]. Clin Chen 1989;35:1121.

 Shah VD. Immunochemical test for CK-MB isoforms [Letter] Clin Chem 1989:35:2157-8.

12. Thompson RJ, Jackson AP, Langlois N. Circulating antibodie to mouse monoclonal immunoglobulins on normal subjects-inci dence, species specificity, and effects in a two-site assay for creating kinase-MB isoenzyme. Clin Chem 1986;32:476-81.

13. Butch AW, Goodnow TT, Brown WS, McClellan A, Kessler G Scott MG. Stratus automated creatine kinase-MB assay evalu ated: identification and elimination of falsely increased result associated with a high molecular mass form of alkaline phospha tase. Clin Chem 1989;35:2048-53.

14. Diamandis EP. Immunoassays with time-resolved fluores cence spectroscopy: principles and applications [Review]. Clin Biochem 1988;21:139-50.

15. Morton RC, Diamandis EP. A streptavidin-based macromolec ular complex labeled with an europium chelator suitable for time-resolved fluoroimmunoassay applications. Anal Chem 1990 in press.

16. Evangelista RA, Pollak A, Allore B, Templeton EF, Morton

RE, Diamandis EP. A new europium chelate for protein labeling and time-resolved fluorometric applications. Clin Biochem 1988;21:73-8.

17. Chan MA, Bellem AC, Diamandis EP. Time-resolved immunofluorometric assay for alpha-fetoprotein in serum and amniotic fluid with a novel detection system. Clin Chem 1987;33:2000-3.

18. Diamandis EP, Morton RC, Reichstein E, Khosravi MJ. Multiple fluorescence labeling with europium chelators. Application to time-resolved fluoroimmunoassays. Anal Chem 1988;61:48–53.

19. Wu AHB, Gornet TG, Harker CC, Chen HL. Role of rapid immunoassays for urgent ("stat") determinations of creatine kinase isoenzyme MB. Clin Chem 1989;35:1752-6.

20. Chapelle JP, Allaf ME. Automated quantification of creatine kinase MB isoenzyme in serum by radial partition immunoassay, with use of the Stratus analyzer. Clin Chem 1990;36:99–101.

21. Brandt DR, Gates RC, Eng KK, et al. Quantifying the MB

isoenzyme of creatine kinase with the Abbott "IMx" immunoassay analyzer. Clin Chem 1990;36:375-8.

 Éisenberg PR, Shaw D, Schaab C, Jaffe AS. Concordance of creatine kinase-MB activity and mass. Clin Chem 1989;35:440-3.
Chan DW, Taylor E, Frye R, Blitzer R. Immunoenzymometric assay for creatine kinase MB with subunit-specific monoclonal antibodies compared with an immunochemical method and electrophoresis. Clin Chem 1985;31:465-9.

24. Apple F, Preese L, Bennett R, Fredrickson A. Clinical and analytical evaluation of two immunoassays for direct measurement of creatine kinase MB with monoclonal anti-CK-MB antibodies. Clin Chem 1988;34:2364-7.

25. Delanghe JR, De Mol AM, Buyzere ML, De Scheerder IK, Wieme RJ. Mass concentration and activity concentration of creatine kinase isoenzyme MB compared in serum after acute myocardial infarction. Clin Chem 1990;36:149–53.

CLIN. CHEM. 36/9, 1683-1685 (1990)

Increase in Lactate Dehydrogenase Isoenzyme-1 in Plasma in Pediatric Malignancy

S. S. Sacks and G. Mulligan

We investigated 28 cases of pediatric malignancy in which total lactate dehydrogenase (LD, EC 1.1.1.27) activities were increased and isoenzyme LD-1 exceeded LD-2 (flipped pattern). Of these, 11 had a flipped pattern at presentation and 17 showed a flipped pattern during chemotherapy. Those with flipped patterns at presentation were four with germ-cell tumors, one with acute lymphocytic leukemia, and six with nephroblastomas (Wilms tumor). Four of five nephroblastoma homogenates contained predominantly LD-1; one revealed a structurally normal LD-1 with normal kinetics. We conclude that an increase in LD with a flipped pattern is pommon in nephroblastoma and, in addition, may develop in cancer patients treated with chemotherapy.

An increase in lactate dehydrogenase (LD, EC 1.1.1.27) in serum is frequently encountered in malignancy. Analysis for LD isoenzymes in these sera may be helpful in distinguishing malignancy from other pathologies, the most frequently encountered malignant patterns being increases of LD-2 and LD-3 (1). Less frequently encountered is a predominant increase of LD-1, so that LD1 > LD2 ("flipped pattern"). This pattern has been described in germ-cell tumors, including seminoma (2) and yolk sac umor (3), and in small cell carcinoma (4). To our knowledge, this occurrence in nephroblastoma has not been previously documented. Here we analyze 28 sequential patients with flipped patterns associated with malignancy.

Materials and Methods

Patients

We analyzed heparinized plasma samples from pediatric patients, ages 12 years and younger, in the Red Cross Children's Hospital. These included 52 patients with nephroblastoma and 24 with germ-cell tumors. Total LD was measured in all samples. All diagnoses of malignancy were histologically confirmed. Staging of nephroblastoma was as follows: stage 1, confined to the kidney; stage 2, spread through the renal capsule; stage 3, involvement of adjacent structures; and stage 4, widespread metastases.

Procedures

Assay of total LD activity. Total LD was measured with a kit (Boehringer Mannheim GmbH) based on a method that conformed to the recommendations of the Deutsche Gesellschaft für Klinische Chemie (5, 6). Assays were run on a centrifugal analyzer (Multistat III Plus; Instrumentation Laboratory, Lexington, MA).

Isoenzyme analysis. All patients' samples with increased total LD were routinely submitted for LD isoenzyme analysis. Isoenzymes of LD were electrophoretically separated with use of the Paragon System and LD isoenzyme electrophoresis kit (Beckman Instruments Inc., Fullerton, CA 92634). Densitometric scans were performed with a scanner (Beckman Computing Densitometer CDS-100F) at 600 nm. Samples with the peak area of LD-1 greater than that of LD-2 were classed as having flipped patterns.

Preparation of tumor homogenate supernates. Nephroblastoma tissue obtained from laparotomy was frozen in liquid nitrogen and stored at -70 °C. A 1-g sample of nephroblastoma was washed in sodium phosphate buffer (50 mmol/L, pH 7.4) and homogenized in 9 mL of buffer with a Thomas glass homogenizer (Thomas Scientific, Swedesboro, NJ). The homogenate was sonicated on ice with a sonicator (Branson Sonic Power Co., Danbury, CT) at 50 W with five bursts of 5 s each, interrupted by 1-min cooling periods. The sonicate was centrifuged for 10 min at $5000 \times g$. Total LD activity in the supernate was measured, adjusted to 1000 U/L with a phosphate buffer, and then isoenzyme activity was determined as described above.

Cardiac muscle and liver tissue were obtained at autopsy 8 h after death, and homogenate supernates were prepared as described for the nephroblastomas.

Department of Chemical Pathology, Groote Schuur Hospital, Cape Town, South Africa.

Received May 29, 1990; accepted June 29, 1990.